

NitraCare Aloe



Powder-Free Nitrile Exam Gloves

Medgluv new Aloe Nitrile Exam Gloves have excellent barrier properties combined with great fit and feel to minimize skin irritation. They offer a soft and comfortable feel with outstanding tactile sensitivity. Naturally helps moisturize and sooth dry, chapped hands.

Description:

- Textured fingertips enables grip, wet or dry.
- Top quality Latex Exam glove with the benefits of aloe vera.
- Naturally helps moisturize and soothe dry, chapped hands.
- Offer excellent sensitivity for touch perception and dependable protection.
- Meets and exceeds NIOSH recommendations for Fentanyl
- Meets both ASTM D6978-05 and USP 800 Personal Protection Equipment requirement for use of select chemotherapy gloves.

Item	Description	Size	Packaging
MG5050A	NitraCare Aloe Nitrile Exam Glove	X-Small	200 eaches/box 10 boxes/case
MG5051A	NitraCare Aloe Nitrile Exam Glove	Small	200 eaches/box 10 boxes/case
MG5052A	NitraCare Aloe Nitrile Exam Glove	Medium	200 eaches/box 10 boxes/case
MG5053A	NitraCare Aloe Nitrile Exam Glove	Large	200 eaches/box 10 boxes/case
MG5054AL	NitraCare Aloe Nitrile Exam Glove	X-Large	200 eaches/box 10 boxes/case

Product Specifications

Gauge Thickness	MM	MIL
Middle Finger:	.09	3.5
Palm:	.08	3.1
Cuff:	.06	2.4
Average Length	242mm	9.5"


Physical Properties

Before Aging

Tensile Strength:	35 MPa
Ultimate Elongation:	580%

After Aging

Tensile Strength:	38 MPa
Ultimate Elongation:	550%

 **Tested with 30 CHEMO DRUGS**
in accordance with ASTM D6978 and barrier tested against Fentanyl exposure.

For additional information, please visit:
www.medgluv.com or call 1-866-633-4588

EXCELLENT BARRIER PROTECTION FOR ALL RISKS



Quality Standards

- Meet and exceeds ASTM D6319 - Standard Specification for Nitrile Examination Gloves for Medical Applications
- Meets and exceeds ASTM F1671 - 07 Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood Borne Pathogens.
- Meets and exceeds ASTM F739-20 Standard Test Method for Permeation of Liquids and Gases Through Protective Clothing Materials Under Conditions of Continuous Contact.
- Meets and exceeds ASTM D6978 Standard Practice For Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs.
- Meet and exceeds ASTM D5151-19 Standard Test for Detection of Holes in Medical Gloves

Specifications are subject to change without notice.

Keep out of sunlight. Store in a cool dry place. Keep away from sources of ozone and ignition.



www.medgluv.com

