

## Powder-Free Nitrile Exam Gloves

The Amethyst dark blue glove is a high performing, latex-free exam glove for use in every application in every environment. Exceptionally stretchy soft, low modulus nitrile synthetic compound excellent for extended periods. Slightly tacky exterior finish provides excellent gripping properties.

### Description:

- Amethyst nitrile exam gloves are strong, highly visible gloves provide the advanced protection and optimal grip needed for demanding, high-stress situations.
- Powder-Free nitrile gloves offer dependable strength and are fully textured for an enhanced grip in demanding environments.
- Meets and exceeds NIOSH recommendations for Fentanyl.
- Meets both ASTM D6978-05 and USP 800 Personal Protection Equipment requirement for use of chemotherapy gloves when handling hazardous drugs.

Item	Description	Size	Packaging
MG5800	Amethyst 8.0 mil Nitrile Exam Glove	X-Small	100 eaches/box 10 boxes/case
MG5801	Amethyst 8.0 mil Nitrile Exam Glove	Small	100 eaches/box 10 boxes/case
MG5802	Amethyst 8.0 mil Nitrile Exam Glove	Medium	100 eaches/box 10 boxes/case
MG5803	Amethyst 8.0 mil Nitrile Exam Glove	Large	100 eaches/box 10 boxes/case
MG5804	Amethyst 8.0 mil Nitrile Exam Glove	X-Large	100 eaches/box 10 boxes/case

## Product Specifications

Gauge Thickness	MM	MIL
Middle Finger:	.205	8.0
Palm:	.16	7.1
Cuff:	.12	6.2
Average Length	242mm	9.5"

## Physical Properties

### Before Aging

Tensile Strength:	35 MPa
Ultimate Elongation:	600%

### After Aging

Tensile Strength:	33 MPa
Ultimate Elongation:	560%



### Tested with 30 CHEMO DRUGS

in accordance with ASTM D6978 and barrier tested against Fentanyl exposure.



This glove meets **NFPA 1999**, Standard on Protective Clothing and Ensembles for Emergency Medical Operations.

For additional information, please visit: [www.medgluv.com](http://www.medgluv.com) or call 1-866-633-4588



## MAXIMUM BARRIER PROTECTION FOR ALL RISKS



view this product  
on our website:



## Quality Standards

- Meet and exceeds ASTM D6319 - Standard Specification for Nitrile Examination Gloves for Medical Applications
- Meets and exceeds ASTM F1671 - 07 Standard Test Method for Resistance of Materials Used in Protective Clothing to Penetration by Blood Borne Pathogens.
- Meets and exceeds ASTM F739-20 Standard Test Method for Permeation of Liquids and Gases Through Protective Clothing Materials Under Conditions of Continuous Contact.
- Meets and exceeds ASTM D6978 Standard Practice For Assessment of Resistance of Medical Gloves to Permeation by Chemotherapy Drugs.
- Meet and exceeds ASTM D5151-19 Standard Test for Detection of Holes in Medical Gloves

Specifications are subject to change without notice.

Keep out of sunlight. Store in a cool dry place. Keep away from sources of ozone and ignition.

